

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00459	Calib. Date: 03/07/2012
Certification File No.:	00460	Calib. No.: 00010
Linearity File No.:	00461	Cert. Date: 03/07/2012
Solution File No.:	00463	Cert. No.: 00008
Sequential File No.:	00463	Lin. Date: 03/07/2012
		Lin. No.: 00008
		Soln. Date: 03/16/2012
		Soln. No.: 00085
		File Date: 03/16/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	11C088	Expires: 03/31/2013
		Bottle No.: 0507

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:50S		
Control 1 EC	0.100%	09:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:51S		
Control 2 EC	0.097%	09:52S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:53S		
Control 3 EC	0.098%	09:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:54S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: 

Badge No.: 129

Date: 03/16/2012