

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00459 Calib. Date: 03/07/2012 Calib. No.: 00010
Certification File No.: 00460 Cert. Date: 03/07/2012 Cert. No.: 00008
Linearity File No.: 00461 Lin. Date: 03/07/2012 Lin. No.: 00008
Solution File No.: 00471 Soln. Date: 04/19/2012 Soln. No.: 00087
Sequential File No.: 00471 File Date: 04/19/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 03/31/2013
Solution Control Lot: 11C088 Bottle No.: 0512

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:17D		
Control 1 EC	0.101%	09:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:18D		
Control 2 EC	0.099%	09:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:20D		
Control 3 EC	0.099%	09:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL #129

Badge No.: 129

Date: 04/19/2012