

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00459	Calib. Date: 03/07/2012
Certification File No.:	00460	Calib. No.: 00010
Linearity File No.:	00461	Cert. Date: 03/07/2012
Solution File No.:	00474	Cert. No.: 00008
Sequential File No.:	00474	Lin. Date: 03/07/2012
		Lin. No.: 00008
		Soln. Date: 05/01/2012
		Soln. No.: 00088
		File Date: 05/01/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	11H094	Expires: 08/17/2013
		Bottle No.: 0321

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:10D		
Control 1 EC	0.101%	13:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:11D		
Control 2 EC	0.098%	13:12D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:12D		
Control 3 EC	0.098%	13:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_



Badge No.: 129

Date: 05/01/2012