

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 00555 Calib. Date: 03/07/2012 Calib. No.: 00015  
Certification File No.: 00556 Cert. Date: 03/07/2012 Cert. No.: 00009  
Linearity File No.: 00557 Lin. Date: 03/07/2012 Lin. No.: 00009  
Solution File No.: 00589 Soln. Date: 05/14/2012 Soln. No.: 00115  
Sequential File No.: 00589 File Date: 05/14/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113  
Control Solution %: 0.100% Expires: 03/13/2013  
Solution Control Lot: 11C088 Bottle No.: 0513

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:59D		
Control 1 EC	0.100%	13:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:00D		
Control 2 EC	0.100%	14:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:01D		
Control 3 EC	0.100%	14:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: COSTANTINI

First Name: JASON

MI: R

Signature: \_\_\_\_\_

Badge No.: 152

Date: 05/14/2012