

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00459 Calib. Date: 03/07/2012 Calib. No.: 00010
Certification File No.: 00460 Cert. Date: 03/07/2012 Cert. No.: 00008
Linearity File No.: 00461 Lin. Date: 03/07/2012 Lin. No.: 00008
Solution File No.: 00487 Soln. Date: 06/04/2012 Soln. No.: 00090
Sequential File No.: 00487 File Date: 06/04/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 08/17/2013
Solution Control Lot: 11H094 Bottle No.: 0316

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:43D		
Control 1 EC	0.101%	22:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:44D		
Control 2 EC	0.098%	22:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:46D		
Control 3 EC	0.098%	22:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLTER

First Name: STEVEN

MI: R

Signature: _____

SGT. Steven Wolter #114

Badge No.: 114

Date: 06/04/2012