

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00555	Calib. Date: 03/07/2012
Certification File No.:	00556	Cert. No.: 00015
Linearity File No.:	00557	Cert. Date: 03/07/2012
Solution File No.:	00624	Lin. No.: 00009
Sequential File No.:	00624	Lin. Date: 03/07/2012
		Soln. No.: 00117
		Soln. Date: 06/24/2012
		File Date: 06/24/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	11C088	Expires: 03/31/2013
		Bottle No.: 0511

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:07D		
Control 1 EC	0.101%	09:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:08D		
Control 2 EC	0.100%	09:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:09D		
Control 3 EC	0.100%	09:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WOLTER

First Name: STEVEN

MI: -

Signature: \_\_\_\_\_

SGT. Steven Wolter #114

Badge No.: 114

Date: 06/24/2012