

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00555 Calib. Date: 03/07/2012 Calib. No.: 00015
Certification File No.: 00556 Cert. Date: 03/07/2012 Cert. No.: 00009
Linearity File No.: 00557 Lin. Date: 03/07/2012 Lin. No.: 00009
Solution File No.: 00646 Soln. Date: 08/01/2012 Soln. No.: 00119
Sequential File No.: 00646 File Date: 08/01/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 08/17/2013
Solution Control Lot: 11H094 Bottle No.: 0261

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:16D		
Control 1 EC	0.099%	12:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:18D		
Control 2 EC	0.099%	12:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:19D		
Control 3 EC	0.099%	12:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____



Badge No.: 162

Date: 08/01/2012