

Calibrating Unit New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARZL-0146 |
| Location: | WALL TOWNSHIP POLICE | |
| Calibration File No.: | 00459 | Calib. Date: 03/07/2012 |
| Certification File No.: | 00460 | Cert. No.: 00010 |
| Linearity File No.: | 00461 | Cert. Date: 03/07/2012 |
| Solution File No.: | 00510 | Lin. Date: 03/07/2012 |
| Sequential File No.: | 00510 | Lin. No.: 00008 |
| | | Soln. Date: 08/01/2012 |
| | | Soln. No.: 00093 |
| | | File Date: 08/01/2012 |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDXB S3-114 |
| Solution Control Lot: | 11H094 | Expires: 08/17/2013 |
| | | Bottle No.: 0252 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 12:15D | | *** TEST PASSED *** |
| Control 1 EC | 0.100% | 12:15D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 12:15D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:16D | | *** TEST PASSED *** |
| Control 2 EC | 0.099% | 12:17D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 12:17D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:17D | | *** TEST PASSED *** |
| Control 3 EC | 0.099% | 12:18D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 12:18D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:19D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 08/01/2012