

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00555	Calib. Date: 03/07/2012
Certification File No.:	00556	Calib. No.: 00015
Linearity File No.:	00557	Cert. Date: 03/07/2012
Solution File No.:	00661	Cert. No.: 00009
Sequential File No.:	00661	Lin. Date: 03/07/2012
		Lin. No.: 00009
		Soln. Date: 08/17/2012
		Soln. No.: 00120
		File Date: 08/17/2012


Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXB S3-113
Control Solution %:	0.100%		Expires: 08/17/2013
Solution Control Lot:	11H094		Bottle No.: 0268

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:05D		
Control 1 EC	0.099%	22:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:06D		
Control 2 EC	0.099%	22:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:07D		
Control 3 EC	0.099%	22:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY	First Name: PTL	MI: A
Signature: 	Badge No.: 129	Date: 08/17/2012