

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00555	Calib. Date:	03/07/2012
Certification File No.:	00556	Cert. Date:	03/07/2012
Linearity File No.:	00557	Lin. Date:	03/07/2012
Solution File No.:	00673	Soln. Date:	08/26/2012
Sequential File No.:	00673	File Date:	08/26/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	11H094	Expires:	08/17/2013
		Bottle No.:	0267

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:51D		
Control 1 EC	0.099%	21:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:53D		
Control 2 EC	0.098%	21:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:54D		
Control 3 EC	0.099%	21:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____



Badge No.: 162

Date: 08/26/2012