

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00681	Calib. Date:	09/04/2012
Certification File No.:	00682	Cert. Date:	09/04/2012
Linearity File No.:	00683	Lin. Date:	09/04/2012
Solution File No.:	00696	Soln. Date:	09/18/2012
Sequential File No.:	00696	File Date:	09/18/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	11H094	Expires:	08/17/2013
		Bottle No.:	0259

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:29D		
Control 1 EC	-.---%	14:29D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	14:29D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PK #129

Badge No.: 129

Date: 09/18/2012