

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE	Calib. No.:	00011
Calibration File No.:	00527	Cert. Date:	09/04/2012
Certification File No.:	00528	Lin. Date:	09/04/2012
Linearity File No.:	00529	Soln. Date:	09/18/2012
Solution File No.:	00534	File Date:	09/18/2012
Sequential File No.:	00534		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAB-0015
Control Solution %:	0.100%			Expires:	08/17/2013
Solution Control Lot:	11H094			Bottle No.:	0275

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:31D		
Control 1 EC	0.097%	16:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.095%	16:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:33D		
Control 2 EC	0.094%	16:34D	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.096%	16:34D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	16:34D		
Control 3 EC	0.094%	16:35D	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.096%	16:35D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	16:36D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 09/18/2012