

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00681	Calib. Date:	09/04/2012
Certification File No.:	00682	Cert. No.:	00016
Linearity File No.:	00683	Cert. Date:	09/04/2012
Solution File No.:	00698	Lin. No.:	00010
Sequential File No.:	00698	Soln. Date:	09/19/2012
		Soln. No.:	00125
		File Date:	09/19/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	11H094	Expires:	08/17/2013
		Bottle No.:	0259

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:43D		
Control 1 EC	0.101%	10:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:44D		
Control 2 EC	0.099%	10:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:45D		
Control 3 EC	0.099%	10:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 09/19/2012