

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00681	Calib. Date:	09/04/2012
Certification File No.:	00682	Cert. Date:	09/04/2012
Linearity File No.:	00683	Lin. Date:	09/04/2012
Solution File No.:	00704	Soln. Date:	10/01/2012
Sequential File No.:	00704	File Date:	10/01/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	11H094	Expires:	08/17/2013
		Bottle No.:	0256

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:56D		
Control 1 EC	0.099%	16:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:57D		
Control 2 EC	0.098%	16:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:59D		
Control 3 EC	0.099%	16:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL [Signature]

Badge No.: 129

Date: 10/01/2012