

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00681	Calib. Date:	09/04/2012
Certification File No.:	00682	Cert. No.:	00016
Linearity File No.:	00683	Cert. Date:	09/04/2012
Solution File No.:	00717	Lin. No.:	00010
Sequential File No.:	00717	Soln. Date:	11/24/2012
		Soln. No.:	00128
		File Date:	11/24/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	11H094	Expires:	08/17/2013
		Bottle No.:	0274

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:47S		
Control 1 EC	0.100%	08:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48S		
Control 2 EC	0.099%	08:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:49S		
Control 3 EC	0.100%	08:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: \_\_\_\_\_

*PTL [Signature]* 105

Badge No.: 165

Date: 11/24/2012

**Received**

NOV 27 2012

e-mailed to #157