

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00527 Calib. Date: 09/04/2012 Calib. No.: 00011
Certification File No.: 00528 Cert. Date: 09/04/2012 Cert. No.: 00009
Linearity File No.: 00529 Lin. Date: 09/04/2012 Lin. No.: 00009
Solution File No.: 00543 Soln. Date: 11/24/2012 Soln. No.: 00101
Sequential File No.: 00543 File Date: 11/24/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 08/17/2013
Solution Control Lot: 11H094 Bottle No.: 0270

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:45S		
Control 1 EC	0.101%	08:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:47S		
Control 2 EC	0.099%	08:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48S		
Control 3 EC	0.099%	08:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: _____

PTC Daniel J. Santoro 165

Badge No.: 165

Date: 11/24/2012

Received

NOV 27 2012

Re-mailed to #157