

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 00681 Calib. Date: 09/04/2012 Calib. No.: 00016  
Certification File No.: 00682 Cert. Date: 09/04/2012 Cert. No.: 00010  
Linearity File No.: 00683 Lin. Date: 09/04/2012 Lin. No.: 00010  
Solution File No.: 00739 Soln. Date: 01/20/2013 Soln. No.: 00130  
Sequential File No.: 00739 File Date: 01/20/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014  
Control Solution %: 0.100% Expires: 08/17/2013  
Solution Control Lot: 11H094 Bottle No.: 0311

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:42S		
Control 1 EC	0.101%	07:43S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:43S		
Control 2 EC	0.100%	07:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:45S		
Control 3 EC	0.100%	07:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: \_\_\_\_\_



Badge No.: 162

Date: 01/20/2013