

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00527	Calib. Date:	09/04/2012
Certification File No.:	00528	Cert. No.:	00011
Linearity File No.:	00529	Cert. Date:	09/04/2012
Solution File No.:	00552	Lin. No.:	00009
Sequential File No.:	00552	Lin. Date:	09/04/2012
		Soln. Date:	01/20/2013
		Soln. No.:	00103
		File Date:	01/20/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	11H094	Expires:	08/17/2013
		Bottle No.:	0324

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:41S		
Control 1 EC	0.101%	07:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:43S		
Control 2 EC	0.099%	07:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:44S		
Control 3 EC	0.099%	07:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: \_\_\_\_\_

Badge No.: 162

Date: 01/20/2013