

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00681	Calib. Date:	09/04/2012
Certification File No.:	00682	Cert. Date:	09/04/2012
Linearity File No.:	00683	Lin. Date:	09/04/2012
Solution File No.:	00742	Soln. Date:	02/01/2013
Sequential File No.:	00742	File Date:	02/01/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	11H094	Expires:	08/17/2013
		Bottle No.:	0262

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:39S		
Control 1 EC	0.101%	11:40S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:40S		
Control 2 EC	0.099%	11:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:42S		
Control 3 EC	0.100%	11:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL K #129

Badge No.: 129

Date: 02/01/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00527	Calib. Date:	09/04/2012
Certification File No.:	00528	Cert. Date:	09/04/2012
Linearity File No.:	00529	Lin. Date:	09/04/2012
Solution File No.:	00554	Soln. Date:	02/01/2013
Sequential File No.:	00554	File Date:	02/01/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	11H094	Expires:	08/17/2013
		Bottle No.:	0325

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:39S		
Control 1 EC	0.101%	11:40S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:41S		
Control 2 EC	0.097%	11:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43S		
Control 3 EC	0.098%	11:43S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

DK. K #129

Badge No.: 129

Date: 02/01/2013