

Calibrating Unit New Standard Solution Report

| | | | |
|-------------------------|-----------------------|--------------|------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARZL-0146 |
| Location: | WALL TOWNSHIP POLICE | | |
| Calibration File No.: | 00527 | Calib. Date: | 09/04/2012 |
| Certification File No.: | 00528 | Cert. No.: | 00011 |
| Linearity File No.: | 00529 | Cert. Date: | 09/04/2012 |
| Solution File No.: | 00554 | Lin. No.: | 00009 |
| Sequential File No.: | 00554 | Lin. Date: | 09/04/2012 |
| | | Soln. No.: | 00009 |
| | | Soln. Date: | 02/01/2013 |
| | | File Date: | 02/01/2013 |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDAB-0015 |
| Solution Control Lot: | 11H094 | Expires: | 08/17/2013 |
| | | Bottle No.: | 0325 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 11:39S | | |
| Control 1 EC | 0.101% | 11:40S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 11:40S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:41S | | |
| Control 2 EC | 0.097% | 11:42S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 11:42S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:43S | | |
| Control 3 EC | 0.098% | 11:43S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 11:43S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:44S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

DK. K #129

Badge No.: 129

Date: 02/01/2013

received
2-1-13

e-mailed to #157