

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00681	Calib. Date:	09/04/2012
Certification File No.:	00682	Cert. Date:	09/04/2012
Linearity File No.:	00683	Lin. Date:	09/04/2012
Solution File No.:	00745	Soln. Date:	02/14/2013
Sequential File No.:	00745	File Date:	02/14/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	11H094	Expires:	08/17/2013
		Bottle No.:	0313

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:52S		
Control 1 EC	0.101%	08:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:53S		
Control 2 EC	0.099%	08:54S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55S		
Control 3 EC	0.098%	08:55S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:56S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL. K #129

Badge No.: 129

Date: 02/14/2013