

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00527	Calib. Date:	09/04/2012
Certification File No.:	00528	Cert. No.:	00011
Linearity File No.:	00529	Cert. Date:	09/04/2012
Solution File No.:	00557	Lin. No.:	00009
Sequential File No.:	00557	Lin. Date:	09/04/2012
		Soln. No.:	00105
		Soln. Date:	02/14/2013
		File Date:	02/14/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	11H094	Expires:	08/17/2013
		Bottle No.:	0319

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:55S		
Control 1 EC	0.101%	08:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:56S		
Control 2 EC	0.097%	08:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:58S		
Control 3 EC	0.098%	08:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*PTL. K #129*

Badge No.: 129

Date: 02/14/2013