

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00562	Calib. Date:	03/01/2013
Certification File No.:	00563	Cert. Date:	03/01/2013
Linearity File No.:	00564	Lin. Date:	03/01/2013
Solution File No.:	00571	Soln. Date:	03/15/2013
Sequential File No.:	00571	File Date:	03/15/2013

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXB S3-113
Control Solution %:	0.100%			Expires:	08/17/2013
Solution Control Lot:	11H094			Bottle No.:	0306

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:06S		
Control 1 EC	0.100%	09:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:07S		
Control 2 EC	0.096%	09:08S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:09S		
Control 3 EC	0.097%	09:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL K #129

Badge No.: 129

Date: 03/15/2013