

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00562	Calib. Date:	03/01/2013
Certification File No.:	00563	Cert. Date:	03/01/2013
Linearity File No.:	00564	Lin. Date:	03/01/2013
Solution File No.:	00575	Soln. Date:	03/17/2013
Sequential File No.:	00575	File Date:	03/17/2013

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXB S3-113
Control Solution %:	0.100%			Expires:	08/17/2013
Solution Control Lot:	11H094			Bottle No.:	0308

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:43S		
Control 1 EC	0.099%	05:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	05:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:44S		
Control 2 EC	0.096%	05:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	05:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:46S		
Control 3 EC	0.096%	05:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 03/17/2013