

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00756	Calib. Date: 03/06/2013
Certification File No.:	00757	Calib. No.: 00019
Linearity File No.:	00758	Cert. Date: 03/06/2013
Solution File No.:	00768	Cert. No.: 00012
Sequential File No.:	00768	Lin. Date: 03/06/2013
		Lin. No.: 00012
		Soln. Date: 03/26/2013
		Soln. No.: 00137
		File Date: 03/26/2013

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	11H094	Expires: 08/17/2013
		Bottle No.: 0302

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:26S		
Control 1 EC	0.101%	15:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:28S		
Control 2 EC	0.101%	15:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:29S		
Control 3 EC	0.100%	15:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:30S	34.0°C	*** TEST PASSED ***
Ambient Air-Blank	0.000%	15:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*PTL. K #129*

Badge No.: 129

Date: 03/26/2013