

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|-----------------------|--------------|-------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARXA-0037 |
| Location: | WALL TOWNSHIP POLICE | | |
| Calibration File No.: | 00756 | Calib. Date: | 03/06/2013 |
| Certification File No.: | 00757 | Cert. Date: | 03/06/2013 |
| Linearity File No.: | 00758 | Lin. Date: | 03/06/2013 |
| Solution File No.: | 00779 | Soln. Date: | 04/12/2013 |
| Sequential File No.: | 00779 | File Date: | 04/12/2013 |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDXB S3-114 |
| Solution Control Lot: | 11H094 | Expires: | 08/17/2013 |
| | | Bottle No.: | 0301 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 08:25D | | |
| Control 1 EC | 0.099% | 08:26D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 08:26D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:26D | | |
| Control 2 EC | 0.098% | 08:27D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 08:27D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:28D | | |
| Control 3 EC | 0.099% | 08:28D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 08:28D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:29D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 04/12/2013