

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 00756 Calib. Date: 03/06/2013 Calib. No.: 00019  
Certification File No.: 00757 Cert. Date: 03/06/2013 Cert. No.: 00012  
Linearity File No.: 00758 Lin. Date: 03/06/2013 Lin. No.: 00012  
Solution File No.: 00787 Soln. Date: 04/29/2013 Soln. No.: 00139  
Sequential File No.: 00787 File Date: 04/29/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114  
Control Solution %: 0.100% Expires: 08/17/2013  
Solution Control Lot: 11H094 Bottle No.: 0273

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 11:11D        |                               |                           |
| Control 1 EC      | 0.099%         | 11:11D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.099%         | 11:11D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:12D        |                               |                           |
| Control 2 EC      | 0.099%         | 11:13D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.099%         | 11:13D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:13D        |                               |                           |
| Control 3 EC      | 0.098%         | 11:14D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.099%         | 11:14D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:15D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_



Badge No.: 129

Date: 04/29/2013