

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00562	Calib. Date:	03/01/2013
Certification File No.:	00563	Cert. Date:	03/01/2013
Linearity File No.:	00564	Lin. Date:	03/01/2013
Solution File No.:	00588	Soln. Date:	04/29/2013
Sequential File No.:	00588	File Date:	04/29/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	11H094	Expires:	08/17/2013
		Bottle No.:	0310

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:10D		
Control 1 EC	0.099%	13:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	13:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:11D		
Control 2 EC	0.097%	13:12D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:13D		
Control 3 EC	0.097%	13:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	13:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 04/29/2013