

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00756	Calib. Date: 03/06/2013
Certification File No.:	00757	Calib. No.: 00019
Linearity File No.:	00758	Cert. Date: 03/06/2013
Solution File No.:	00798	Cert. No.: 00012
Sequential File No.:	00798	Lin. Date: 03/06/2013
		Lin. No.: 00012
		Soln. Date: 05/17/2013
		Soln. No.: 00140
		File Date: 05/17/2013

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXB S3-114
Control Solution %: 0.100%		Expires: 08/17/2013
Solution Control Lot: 11H094		Bottle No.: 0307

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:45D		
Control 1 EC	0.098%	14:45D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	14:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:46D		
Control 2 EC	0.097%	14:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	14:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:47D		
Control 3 EC	0.097%	14:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*PTL K #129*

Badge No.: 129

Date: 05/17/2013