

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00562	Calib. Date: 03/01/2013
Certification File No.:	00563	Calib. No.: 00012
Linearity File No.:	00564	Cert. Date: 03/01/2013
Solution File No.:	00590	Cert. No.: 00010
Sequential File No.:	00590	Lin. Date: 03/01/2013
		Lin. No.: 00010
		Soln. Date: 05/17/2013
		Soln. No.: 00115
		File Date: 05/17/2013

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	11H094	Expires: 08/17/2013
		Bottle No.: 0303

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:57D		
Control 1 EC	0.096%	19:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:58D		
Control 2 EC	0.095%	19:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	19:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:00D		
Control 3 EC	0.096%	20:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 05/17/2013