

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00562	Calib. Date:	03/01/2013
Certification File No.:	00563	Cert. Date:	03/01/2013
Linearity File No.:	00564	Lin. Date:	03/01/2013
Solution File No.:	00593	Soln. Date:	05/23/2013
Sequential File No.:	00593	File Date:	05/23/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0923

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:48D		
Control 1 EC	0.095%	13:49D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	13:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:50D		
Control 2 EC	0.094%	13:51D	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.097%	13:51D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	13:51D		
Control 3 EC	0.094%	13:52D	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.098%	13:52D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	13:53D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PK. K #129

Badge No.: 129

Date: 05/23/2013