

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00562	Calib. Date: 03/01/2013
Certification File No.:	00563	Calib. No.: 00012
Linearity File No.:	00564	Cert. Date: 03/01/2013
Solution File No.:	00594	Cert. No.: 00010
Sequential File No.:	00594	Lin. Date: 03/01/2013
		Lin. No.: 00010
		Soln. Date: 05/23/2013
		Soln. No.: 00117
		File Date: 05/23/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	12F102	Expires: 06/28/2014
		Bottle No.: 0923

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:30D		
Control 1 EC	0.096%	16:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:32D		
Control 2 EC	0.095%	16:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	16:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:33D		
Control 3 EC	0.096%	16:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	16:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PK. K #109

Badge No.: 129

Date: 05/23/2013