

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00756	Calib. Date:	03/06/2013
Certification File No.:	00757	Cert. No.:	00019
Linearity File No.:	00758	Lin. Date:	03/06/2013
Solution File No.:	00809	Cert. No.:	00012
Sequential File No.:	00809	Lin. No.:	00012
		Soln. No.:	00141
		File Date:	05/25/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	11H094	Expires:	08/17/2013
		Bottle No.:	0304

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:28D		
Control 1 EC	0.100%	23:29D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:29D		
Control 2 EC	0.098%	23:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:31D		
Control 3 EC	0.097%	23:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 05/25/2013