

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00562 Calib. Date: 03/01/2013 Calib. No.: 00012
Certification File No.: 00563 Cert. Date: 03/01/2013 Cert. No.: 00010
Linearity File No.: 00564 Lin. Date: 03/01/2013 Lin. No.: 00010
Solution File No.: 00597 Soln. Date: 05/25/2013 Soln. No.: 00118
Sequential File No.: 00597 File Date: 05/25/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 06/28/2014
Solution Control Lot: 12F102 Bottle No.: 0921

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:32D		
Control 1 EC	0.098%	23:33D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:34D		
Control 2 EC	0.096%	23:34D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:35D		
Control 3 EC	0.097%	23:36D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____



Badge No.: 162

Date: 05/25/2013