

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00756	Calib. Date:	03/06/2013
Certification File No.:	00757	Cert. Date:	03/06/2013
Linearity File No.:	00758	Lin. Date:	03/06/2013
Solution File No.:	00820	Soln. Date:	06/13/2013
Sequential File No.:	00820	File Date:	06/13/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0910

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:15D		
Control 1 EC	0.099%	07:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:16D		
Control 2 EC	0.099%	07:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:18D		
Control 3 EC	0.099%	07:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 06/13/2013