

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00562	Calib. Date: 03/01/2013
Certification File No.:	00563	Calib. No.: 00012
Linearity File No.:	00564	Cert. Date: 03/01/2013
Solution File No.:	00605	Cert. No.: 00010
Sequential File No.:	00605	Lin. Date: 03/01/2013
		Lin. No.: 00010
		Soln. Date: 06/13/2013
		Soln. No.: 00119
		File Date: 06/13/2013

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXB S3-113
Control Solution %:	0.100%			Expires:	06/28/2014
Solution Control Lot:	12F102			Bottle No.:	0922

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:15D		
Control 1 EC	0.099%	07:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:16D		
Control 2 EC	0.097%	07:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	07:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:17D		
Control 3 EC	0.097%	07:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 06/13/2013