

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00756	Calib. Date: 03/06/2013
Certification File No.:	00757	Calib. No.: 00019
Linearity File No.:	00758	Cert. Date: 03/06/2013
Solution File No.:	00830	Cert. No.: 00012
Sequential File No.:	00830	Lin. Date: 03/06/2013
		Lin. No.: 00012
		Soln. Date: 06/28/2013
		Soln. No.: 00143
		File Date: 06/28/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	12F102	Expires: 06/28/2014
		Bottle No.: 0901

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:32D		
Control 1 EC	0.101%	17:33D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:33D		
Control 2 EC	0.099%	17:34D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:35D		
Control 3 EC	0.099%	17:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: _____



Badge No.: 165

Date: 06/28/2013