

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00562	Calib. Date:	03/01/2013
Certification File No.:	00563	Cert. Date:	03/01/2013
Linearity File No.:	00564	Lin. Date:	03/01/2013
Solution File No.:	00612	Soln. Date:	06/28/2013
Sequential File No.:	00612	File Date:	06/28/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0918

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:07D		
Control 1 EC	0.099%	17:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:08D		
Control 2 EC	0.097%	17:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	17:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:10D		
Control 3 EC	0.097%	17:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: \_\_\_\_\_



Badge No.: 165

Date: 06/28/2013