

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00756	Calib. Date: 03/06/2013
Certification File No.:	00757	Calib. No.: 00019
Linearity File No.:	00758	Cert. Date: 03/06/2013
Solution File No.:	00842	Cert. No.: 00012
Sequential File No.:	00842	Lin. Date: 03/06/2013
		Lin. No.: 00012
		Soln. Date: 07/08/2013
		Soln. No.: 00144
		File Date: 07/08/2013

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXB S3-114
Control Solution %: 0.100%		Expires: 06/28/2014
Solution Control Lot: 12F102		Bottle No.: 0908

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:30D		
Control 1 EC	0.101%	03:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:32D		
Control 2 EC	0.100%	03:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:33D		
Control 3 EC	0.100%	03:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 07/08/2013