

Calibrating Unit New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARZL-0146 |
| Location: | WALL TOWNSHIP POLICE | |
| Calibration File No.: | 00562 | Calib. Date: 03/01/2013 |
| Certification File No.: | 00563 | Calib. No.: 00012 |
| Linearity File No.: | 00564 | Cert. Date: 03/01/2013 |
| Solution File No.: | 00631 | Cert. No.: 00010 |
| Sequential File No.: | 00631 | Lin. Date: 03/01/2013 |
| | | Lin. No.: 00010 |
| | | Soln. Date: 07/17/2013 |
| | | Soln. No.: 00122 |
| | | File Date: 07/17/2013 |

| | | | | | |
|-----------------------|--------|------------|-------|-------------|-------------|
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDXB S3-113 |
| Control Solution %: | 0.100% | | | Expires: | 06/28/2014 |
| Solution Control Lot: | 12F102 | | | Bottle No.: | 0924 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 02:15D | | |
| Control 1 EC | 0.099% | 02:15D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 02:15D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:16D | | |
| Control 2 EC | 0.097% | 02:17D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 02:17D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:17D | | |
| Control 3 EC | 0.097% | 02:18D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 02:18D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:19D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TOZER

First Name: MICHAEL

MI: R

Signature: _____

Sgt. Muller 132

Badge No.: 132

Date: 07/17/2013