

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00756	Calib. Date:	03/06/2013
Certification File No.:	00757	Cert. No.:	00019
Linearity File No.:	00758	Cert. Date:	03/06/2013
Solution File No.:	00861	Lin. Date:	03/06/2013
Sequential File No.:	00861	Lin. No.:	00012
		Soln. Date:	07/19/2013
		Soln. No.:	00145
		File Date:	07/19/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0907

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:49D		
Control 1 EC	0.100%	09:49D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:50D		
Control 2 EC	0.099%	09:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:51D		
Control 3 EC	0.099%	09:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*[Handwritten Signature]* #129

Badge No.: 129

Date: 07/19/2013

*E-mailed to 157  
7-19-13*