

# Calibrating Unit

## New Standard Solution Report

|                         |                       |              |             |
|-------------------------|-----------------------|--------------|-------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.:  | ARXA-0037   |
| Location:               | WALL TOWNSHIP POLICE  |              |             |
| Calibration File No.:   | 00756                 | Calib. Date: | 03/06/2013  |
| Certification File No.: | 00757                 | Cert. No.:   | 00019       |
| Linearity File No.:     | 00758                 | Cert. Date:  | 03/06/2013  |
| Solution File No.:      | 00872                 | Lin. No.:    | 00012       |
| Sequential File No.:    | 00872                 | Lin. Date:   | 03/06/2013  |
|                         |                       | Soln. No.:   | 00146       |
|                         |                       | Soln. Date:  | 08/01/2013  |
|                         |                       | File Date:   | 08/01/2013  |
| <br>                    |                       |              |             |
| Calibrating Unit:       | WET                   | Model No.:   | CU-34       |
| Control Solution %:     | 0.100%                | Serial No.:  | DDXB S3-114 |
| Solution Control Lot:   | 12F102                | Expires:     | 06/28/2014  |
|                         |                       | Bottle No.:  | 0909        |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 02:44D |                |                     |
| Control 1 EC      | 0.099% | 02:45D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.099% | 02:45D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:46D |                |                     |
| Control 2 EC      | 0.097% | 02:46D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.099% | 02:46D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:47D |                |                     |
| Control 3 EC      | 0.097% | 02:47D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.099% | 02:47D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:48D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TOZER

First Name: MICHAEL

MI: R

Signature: \_\_\_\_\_

Sgt. Murphy 132

Badge No.: 132

Date: 08/01/2013