

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00562	Calib. Date:	03/01/2013
Certification File No.:	00563	Cert. Date:	03/01/2013
Linearity File No.:	00564	Lin. Date:	03/01/2013
Solution File No.:	00640	Soln. Date:	08/01/2013
Sequential File No.:	00640	File Date:	08/01/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0917

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:46D		
Control 1 EC	0.099%	02:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:48D		
Control 2 EC	0.096%	02:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:49D		
Control 3 EC	0.096%	02:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TOZER

First Name: MICHAEL

MI: R

Signature: _____

Sgt. Murphy 132

Badge No.: 132

Date: 08/01/2013