

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00756	Calib. Date: 03/06/2013
Certification File No.:	00757	Calib. No.: 00019
Linearity File No.:	00758	Cert. Date: 03/06/2013
Solution File No.:	00878	Cert. No.: 00012
Sequential File No.:	00878	Lin. Date: 03/06/2013
		Lin. No.: 00012
		Soln. Date: 08/08/2013
		Soln. No.: 00147
		File Date: 08/08/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	12F102	Expires: 06/28/2014
		Bottle No.: 0902

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:53D		
Control 1 EC	0.099%	13:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:54D		
Control 2 EC	0.097%	13:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:55D		
Control 3 EC	0.097%	13:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PK #129

Badge No.: 129

Date: 08/08/2013