

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00756 Calib. Date: 03/06/2013 Calib. No.: 00019
Certification File No.: 00757 Cert. Date: 03/06/2013 Cert. No.: 00012
Linearity File No.: 00758 Lin. Date: 03/06/2013 Lin. No.: 00012
Solution File No.: 00892 Soln. Date: 08/19/2013 Soln. No.: 00148
Sequential File No.: 00892 File Date: 08/19/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 06/28/2014
Solution Control Lot: 12F102 Bottle No.: 0903

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:43D		
Control 1 EC	0.099%	00:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:44D		
Control 2 EC	0.097%	00:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:45D		
Control 3 EC	0.097%	00:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHALLORAN

First Name: SEAN

MI: P

Signature: _____



Badge No.: 146

Date: 08/19/2013

*e-mailed to #157
8-20-13*