

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00756	Calib. Date: 03/06/2013
Certification File No.:	00757	Cert. No.: 00019
Linearity File No.:	00758	Cert. Date: 03/06/2013
Solution File No.:	00895	Lin. Date: 03/06/2013
Sequential File No.:	00895	Lin. No.: 00012
		Soln. Date: 08/24/2013
		Soln. No.: 00149
		File Date: 08/24/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	12F102	Expires: 06/28/2014
		Bottle No.: 0905

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:02D		
Control 1 EC	0.099%	18:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:03D		
Control 2 EC	0.098%	18:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:04D		
Control 3 EC	0.098%	18:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_



Badge No.: 129

Date: 08/24/2013