

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00899	Calib. Date:	08/28/2013
Certification File No.:	00900	Cert. No.:	00020
Linearity File No.:	00901	Cert. Date:	08/28/2013
Solution File No.:	00914	Lin. No.:	00013
Sequential File No.:	00914	Lin. Date:	08/28/2013
		Soln. Date:	09/06/2013
		Soln. No.:	00152
		File Date:	09/06/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0915

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:36D		
Control 1 EC	0.098%	14:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:38D		
Control 2 EC	0.096%	14:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:39D		
Control 3 EC	0.096%	14:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL K #129

Badge No.: 129

Date: 09/06/2013