

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00899	Calib. Date:	08/28/2013
Certification File No.:	00900	Cert. Date:	08/28/2013
Linearity File No.:	00901	Lin. Date:	08/28/2013
Solution File No.:	00913	Soln. Date:	09/06/2013
Sequential File No.:	00913	File Date:	09/06/2013

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAB-0015
Control Solution %:	0.100%			Expires:	06/28/2014
Solution Control Lot:	12F102			Bottle No.:	0914

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:30D		
Control 1 EC	0.099%	04:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	04:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:31D		
Control 2 EC	0.097%	04:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:32D		
Control 3 EC	0.097%	04:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:33D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KELLY

First Name: JASON

MI: R

Signature: 

Badge No.: 166

Date: 09/06/2013