

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00899	Calib. Date: 08/28/2013
Certification File No.:	00900	Cert. No.: 00020
Linearity File No.:	00901	Cert. Date: 08/28/2013
Solution File No.:	00930	Lin. No.: 00013
Sequential File No.:	00930	Lin. Date: 08/28/2013
		Soln. No.: 00154
		Soln. Date: 10/12/2013
		File Date: 10/12/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	13F114	Expires: 06/17/2015
		Bottle No.: 1121

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:14D		
Control 1 EC	0.101%	08:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:15D		
Control 2 EC	0.100%	08:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16D		
Control 3 EC	0.099%	08:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL K #129

Badge No.: 129

Date: 10/12/2013