

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00655	Calib. Date: 08/28/2013
Certification File No.:	00656	Calib. No.: 00013
Linearity File No.:	00657	Cert. Date: 08/28/2013
Solution File No.:	00670	Cert. No.: 00011
Sequential File No.:	00670	Lin. Date: 08/28/2013
		Lin. No.: 00011
		Soln. Date: 10/12/2013
		Soln. No.: 00131
		File Date: 10/12/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	13F114	Expires: 06/17/2015
		Bottle No.: 1112

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:43D		
Control 1 EC	-.---%	07:44D	24.4°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	07:44D	24.4°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL K #129

Badge No.: 129

Date: 10/12/2013